



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/814,979	03/30/2004	Scott Sibbett	21058/0206803-US0	8637
75172	7590	05/01/2009		
Client 21058 c/o DARBY & DARBY P.C. P.O. BOX 770 CHURCH STREET STATION NEW YORK, NY 10008-0770			EXAMINER NOGUEROLA, ALEXANDER STEPHAN	
			ART UNIT 1795	PAPER NUMBER
			MAIL DATE 05/01/2009	DELIVERY MODE PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

**Advisory Action  
Before the Filing of an Appeal Brief**

Application No.

10/814,979

Applicant(s)

SIBBETT, SCOTT

Examiner

ALEX NOGUEROLA

Art Unit

1795

**--The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**

THE REPLY FILED 3 FAILS TO PLACE THIS APPLICATION IN CONDITION FOR ALLOWANCE.

1. ☒ The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:

- a) ☒ The period for reply expires 3 months from the mailing date of the final rejection.  
b) ☐ The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.

Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f).

Extensions of time may be obtained under 37 CFR 1.136(a). The date on which the petition under 37 CFR 1.136(a) and the appropriate extension fee have been filed is the date for purposes of determining the period of extension and the corresponding amount of the fee. The appropriate extension fee under 37 CFR 1.17(a) is calculated from: (1) the expiration date of the shortened statutory period for reply originally set in the final Office action; or (2) as set forth in (b) above, if checked. Any reply received by the Office later than three months after the mailing date of the final rejection, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**NOTICE OF APPEAL**

2. ☐ The Notice of Appeal was filed on \_\_\_\_\_. A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).

**AMENDMENTS**

3. ☒ The proposed amendment(s) filed after a final rejection, but prior to the date of filing a brief, will not be entered because  
(a) ☒ They raise new issues that would require further consideration and/or search (see NOTE below);  
(b) ☒ They raise the issue of new matter (see NOTE below);  
(c) ☒ They are not deemed to place the application in better form for appeal by materially reducing or simplifying the issues for appeal; and/or  
(d) ☐ They present additional claims without canceling a corresponding number of finally rejected claims.

NOTE: See Continuation Sheet. (See 37 CFR 1.116 and 41.33(a)).

4. ☐ The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).  
5. ☐ Applicant's reply has overcome the following rejection(s): \_\_\_\_\_.  
6. ☐ Newly proposed or amended claim(s) \_\_\_\_\_ would be allowable if submitted in a separate, timely filed amendment canceling the non-allowable claim(s).  
7. ☒ For purposes of appeal, the proposed amendment(s): a) ☒ will not be entered, or b) ☐ will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.  
The status of the claim(s) is (or will be) as follows:  
Claim(s) allowed: \_\_\_\_\_.  
Claim(s) objected to: \_\_\_\_\_.  
Claim(s) rejected: 1-3, 5-11, and 13-25.  
Claim(s) withdrawn from consideration: \_\_\_\_\_.

**AFFIDAVIT OR OTHER EVIDENCE**

8. ☐ The affidavit or other evidence filed after a final action, but before or on the date of filing a Notice of Appeal will not be entered because applicant failed to provide a showing of good and sufficient reasons why the affidavit or other evidence is necessary and was not earlier presented. See 37 CFR 1.116(e).  
9. ☐ The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will not be entered because the affidavit or other evidence failed to overcome all rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).  
10. ☐ The affidavit or other evidence is entered. An explanation of the status of the claims after entry is below or attached.

**REQUEST FOR RECONSIDERATION/OTHER**

11. ☐ The request for reconsideration has been considered but does NOT place the application in condition for allowance because: \_\_\_\_\_.  
12. ☐ Note the attached Information *Disclosure Statement*(s). (PTO/SB/08) Paper No(s). \_\_\_\_\_.  
13. ☐ Other: \_\_\_\_\_.

/Alex Noguerola/  
Primary Examiner, Art Unit 1795

Continuation of 3. NOTE: Applicant proposes to amend independent claims 1 and 11 by requiring that "either the first pump channel or the second pump channel is coated to suppress electroosmotic flow." However, claims 1 and 11 already require "a voltage drop between the first and second pump electrodes causes electroosmotic flow in the first and second pump channels ...". If electroosmotic flow is suppressed in the first or second pump channel how can there also be electroosmotic flow in the first and second pump channels? The proposed limitation may also raise a question of new matter as there does not appear to be any mention in the original disclosure of using a coating to suppress electroosmosis. The only reference to a channel coating is in paragraph [0029] of the specification, which discloses using a coating to minimize "absorption of charged molecules." Claim 22 has been apparently inadvertently amended to require the steps of "applying an electric field gradient." There is no support for an "electric field gradient" in the original disclosure. Moreover, as stated in the previous Office action, at the bottom of page 6, "The enablement problems for claims 19-21 similarly apply to claims 22-25." Claim 22 is an independent claim that has not been amended other than as just mentioned nor has its enablement been explained. As for the proposed replacement Figure 4 the Examiner declines to enter it. It is true that paragraph [0024] of the specification discloses reversing electroosmotic flow. However, doing so, so that the electroosmotic flow is from reservoir 116 to reservoir 114 would not create a convective flow in the reverse direction from that shown in the original Figure 4. Indeed, it is not readily apparent that convective flow would occur at all, since reversed electroosmotic flow would create positive pressure, not negative pressure, at the junction 124..